Certificate of Analysis

LIQUILABS
Analytical services

Purexa
Purexa Peptide
Germany
Purexa-Peptide.de

Liquilabs s.r.o.
Inovační 122
252 41 Zlatníky-Hodkovice
Czechia
www.liquilabs.cz

Verify Results Online

Sample Identification


Sample Name	Selank 10 mg
Batch Number	701554
Date Published	2026-09-04 13:17

Results for LYO-0660


Peptides	Result	Unit	Uncertainty	Acceptable Range
Selank Assay Peptide Screening 0.1% TFA	10.87	*mg*	[± 0.05]	
Selank Purity Peptide Screening 0.1% TFA	> 99.8	*%*		
Selank Identification by Spectrum (FTIR) Peptide Screening 0.1% TFA	976		[± 5]	
Selank Identification by RT Peptide Screening 0.1% TFA	0.983		[± 0.005]	

Microbiology	Result	Unit	Uncertainty	Acceptable Range
Bacterial Endotoxin Chromgenic USP<85>/ Eur. Ph. 2.6.14. Bacterial Endotoxin Chromgenic Test	0.034	*EU/mg*	[± 0.001]	

Elemental Impurities	Result	Unit	Uncertainty	Acceptable Range
Arsenic Elemental Impurities screening USP ⟨232⟩ / Ph. Eur. 5.20 / 2.4.20	< 0.001	*ppm*		
Cadmium Elemental Impurities screening USP ⟨232⟩ / Ph. Eur. 5.20 / 2.4.20	< 0.001	*ppm*		
Quicksilver Elemental Impurities screening USP ⟨232⟩ / Ph. Eur. 5.20 / 2.4.20	< 0.001	*ppm*		
Lead Elemental Impurities screening USP ⟨232⟩ / Ph. Eur. 5.20 / 2.4.20	< 0.001	*ppm*		
Nickel Elemental Impurities screening USP ⟨232⟩ / Ph. Eur. 5.20 / 2.4.20	< 0.001	*ppm*		
Vanadium Elemental Impurities screening USP ⟨232⟩ / Ph. Eur. 5.20 / 2.4.20	< 0.001	*ppm*		
Cobalt Elemental Impurities screening USP ⟨232⟩ / Ph. Eur. 5.20 / 2.4.20	< 0.001	*ppm*		



Liquilabs s.r.o. • Inovační 122 • 252 41 Zlatníky-Hodkovice • Czechia
Phone : +420 608 444 727 • Fax : • jan.galbavy@liquilabs.cz • www.liquilabs.cz



Liquilabs s.r.o. • Inovační 122 • 252 41 Zlatníky-Hodkovice • Czechia
Phone : +420 608 444 727 • Fax : • jan.galbavy@liquilabs.cz • www.liquilabs.cz

1. Content Assesment

1.1. Instrumentation

Module	Name	Serial Number
System Controller	Shimadzu CBM-40 Lite	L221226351398
Degassing Unit	Shimadzu DGU-403	NA
Pump	Shimadzu LC-40B XR	L22146350580
Autosampler	Shimadzu SIL-40C XR	L22216351622
Colum Thermostat	Shimadzu CTO-40S	L22236351602
PDA Detector	Shimadzu SPD-M40	L22276352808
SQ MS Detector	Shimadzu LCMS-2050	O12476200760

1.2. Chromatographic conditions

Chromatographic conditions	
Eluent A	0.1% DFA in Water (HPLC, Gradient Grade)
Eluent B	0.1% DFAin Acetonitrile (HPLC, Gradient Grade)
Flow rate	0.8 mL/min
Program	Gradient elution
Injection volume	2 µL
Colum Temperature	60°C
Column	Waters XSelect CSH C18, 100x2.1mm 2.5µm
Detection wavelength	225 nm

Gradient Program		
Time [min]	A [%]	B [%]
1	98	2
13.99	35	65
14	5	95
15	5	95
15.01	98	2
19	end	

1


Liquilabs s.r.o. • Inovační 122 • 252 41 Zlatníky-Hodkovice • Czechia
Phone : +420 608 444 727 • Fax : • jan.galbavy@liquilabs.cz • www.liquilabs.cz

1.3. Sample preparation

Whole amount of container was dissolved in 2mL of water (LCMS Grade). 100 µL of sample was transferred to HPLC vial and diluted by 900 µL water (LCMS Grade) and submitted for analysis.

1.4. Calibration curve

Calibration curve detail	
Quantitative method	External Standard
Calibration Type	Linear
Number of calibration points	5
Force through Zero	Enabled
Weighting Method	None



2

Liquilabs s.r.o. • Inovační 122 • 252 41 Zlatníky-Hodkovice • Czechia
Phone : +420 608 444 727 • Fax : • jan.galbavy@liquilabs.cz • www.liquilabs.cz

2. Purity assessment

2.1 Instrumentation

Module	Name	Serial Number
System Controller	Shimadzu CBM-40 Lite	L221226351398
Degassing Unit	Shimadzu DGU-403	NA
Pump	Shimadzu LC-40B XR	L22146350580
Autosampler	Shimadzu SIL-40C XR	L22216351622
Colum Thermostat	Shimadzu CTO-40S	L22236351602
PDA Detector	Shimadzu SPD-M40	L22276352808
SQ MS Detector	Shimadzu LCMS-2050	O12476200760

2.2 Chromatographic conditions

Chromatographic conditions	
Eluent A	0.1% DFA in Water (HPLC, Gradient Grade)
Eluent B	0.1% DFAin Acetonitrile (HPLC, Gradient Grade)
Flow rate	0.8 mL/min
Program	Gradient elution
Injection volume	2 µL
Colum Temperature	60°C
Column	Waters XSelect CSH C18, 100x2.1mm 2.5µm
Detection wavelength	225nm

Gradient Program		
Time [min]	A [%]	B [%]
1	98	2
13.99	35	65
14	5	95
15	5	95
15.01	98	2
19	end	

2.3 Purity assesment

Purity of compound assesed by area normalization method, comparing area of each peak to sum of area of all peaks detected at wavelenght of 225 nm.

3

Liquilabs s.r.o. • Inovační 122 • 252 41 Zlatníky-Hodkovice • Czechia
Phone : +420 608 444 727 • Fax : • jan.galbavy@liquilabs.cz • www.liquilabs.cz

Analysis Report


LIQUILABS Analytical services


Analysis of quantity and purity of active ingredient by UHPLC with UV detection

Sample Information

Injection Volume	: 2
Data File	: LYO-0660__004.lcd
Method File	: Peptide Screening_DFA_Group E.lcm
Date Acquired	: 8/25/2026 1:31:24 PM

Chromatogram





Peak Table

PDA Ch1 225nm

Name	Ret. Time	Area	Conc.	Unit	Area%
Selank	0.893	419416	10.865	mg	100.000
		419416			100.000

Peak Table

PDA Ch2 280nm

Name	Ret. Time	Area	Conc.	Unit


1 / 1

Liquilabs s.r.o. • Inovační 122 • 252 41 Zlatníky-Hodkovice • Czechia
Phone : +420 608 444 727 • Fax : • jan.galbavy@liquilabs.cz • www.liquilabs.cz

Endotoxin Determination Report

File Information
Filename: C:\UVVis-Data\Data\File_260825_112353.vqud
Date/Time: 08/25/2026 12:17:16 PM

Instrument Information
Instrument Type: UV-1900 Series
Model (S/N): UV-1900i (A12536253123)



y = 20.7814 x + 0.0235192
r2 = 0.99830

	Sample Name	Conc	Raw_WL545.0	Dilution Rate
1	LYO-0660	0.367	0.4049	20.00
2	LYO-0661	0.125	0.1530	20.00
3	LYO-0662	0.084	0.1111	20.00
4	LYO-0663	2.058	1.0927	40.00
5	LYO-0664	0.196	0.2275	20.00
6	LYO-0666	0.101	0.1284	20.00
7	LYO-0667	0.068	0.0942	20.00
8	LYO-0668	0.080	0.1066	20.00
9	LYO-0669	1.030	0.5587	40.00
10	LYO-0672	0.113	0.1411	20.00
11	LYO-0671	0.311	0.1529	50.00
12	LYO-0672	2.514	0.5459	100.00
13	LYO-0673	0.066	0.0925	20.00
14	LYO-0674	2.276	0.6147	80.00
15	LYO-0675	0.074	0.1001	20.00
16	LYO-0676	0.059	0.0845	20.00
17	LYO-0677	0.136	0.1649	20.00
18	LYO-0678	3.182	0.8502	80.00

Liquilabs s.r.o. • Inovační 122 • 252 41 Zlatníky-Hodkovice • Czechia
Phone : +420 608 444 727 • Fax : • jan.galbavy@liquilabs.cz • www.liquilabs.cz

Responsibles



Mr. Ján Galbavý
CEO

Analysis results relate only to the samples tested.
This document shall not be reproduced except in full, without the written approval of Liquilabs s.r.o.

Liquilabs s.r.o. • Inovační 122 • 252 41 Zlatníky-Hodkovice • Czechia
Phone : +420 608 444 727 • Fax : • jan.galbavy@liquilabs.cz • www.liquilabs.cz